

MEASUREMENT PROTOCOL

No. PM_TTP_OPT_RR_MM_DD_XX

Order number 123456
Customer ZČU, TTP

Sample label customer S123456
 NTC 210417-001/002


Quantity Spectral normal hemispherical reflectance absolute
Measurement method SNHRRT/V1 - UV-VIS-NIR
Temperature (T) 25°C
Polar angle (θ) 8°
Azimuth angle (φ) 0°, 45°, 90°, 135°

Date 17.4.2021

Measured by Ilona Slavíková

Responsible person Ing. Petra Honnerová, Ph.D.

Phone: +420 377 634 719
 Email: petrahon@ntc.zcu.cz



WL (um)	0.35	0.40	0.50	0.60	0.80	1.00	1.10	1.20
reflectance (%)	9.5	12.7	21.7	49.0	73.4	75.1	75.3	75.3