


MEASUREMENT PROTOCOL

No. PM_TTP_OPT_RR_MM_DD_XX

Order number		123456	
Customer		ZČU, TTP	
Sample label	customer	S123456	
	NTC	210417-001/003	
Quantity		Spectral normal hemispherical transmittance	
Measurement method		SNHTRT - UV-VIS-NIR	
Temperature (T)		25°C	
Polar angle		0°	
Date		17.4.2021	
Measured by		Ilona Slavíková	
Responsible person		Ing. Petra Honnerová, Ph.D.	Phone: +420 377 634 719 Email: petrahon@ntc.zcu.cz



wavelength (µm)	0.35	0.40	0.50	0.60	0.80	1.00	1.10	1.20
transmittance (%)	1.5	1.4	12.4	5.4	41.5	50.5	50.9	49.7